



基底: 氟化钙

材料等级: 红外多晶

镀膜: 未镀膜

编码	直径 (mm)	直径公差 (mm)	厚度 (mm)	厚度公差 (mm)	材料等级	光洁度	N(ΔN) /PV	平行度 (arcmin)	崩边 (mm)	倒边 (mm)	通光 口径(%)
PCa01B011	6	+0.0/-0.1	1	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01C026	12	+0.0/-0.1	3	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01C027	13	+0.0/-0.1	1	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01C028	13	+0.0/-0.1	2	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01C029	15	+0.0/-0.1	2	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01D032	17.5	+0.0/-0.1	4	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01D033	18	+0.0/-0.1	5	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01D034	19	+0.0/-0.1	2	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01D035	20	+0.0/-0.1	2.5	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01D036	20	+0.0/-0.1	3	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01E077	22	+0.0/-0.1	4	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01E078	25	+0.0/-0.1	4	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01E079	25	+0.0/-0.1	5	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01E080	25.2	+0.0/-0.1	5	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F076	37.5	+0.0/-0.1	4	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F077	38	+0.0/-0.1	3	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F078	38	+0.0/-0.1	6	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F079	41	+0.0/-0.1	3	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F080	49	+0.0/-0.1	6	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95
PCa01F081	50	+0.0/-0.1	3	± 0.1	IR-POLY	60-40	5(3)	<3	0.35	0.35	>95